



23446

PATENT TRADEMARK OFFICE

United States Patent [19]

Bueche et al.



US005851172A

[11] Patent Number: 5,851,172

[45] Date of Patent: Dec. 22, 1998

[54] AFTERLOADER WITH ACTIVE FORCE  
FEEDBACK3643893 6/1988 Germany ..... 600/7  
0649412 3/1979 U.S.S.R. .... 600/3[75] Inventors: Kenneth M. Bueche, Friendswood;  
Richard T. Thornton, League City;  
Anthony J. Bradshaw, Missouri City,  
all of Tex.

Primary Examiner—John P. Lacyk

Attorney, Agent, or Firm—Wigman, Cohen, Leitner &  
Myers, P.C.[73] Assignee: Omnitron International, Inc.,  
Houston, Tex.

[57] ABSTRACT

[21] Appl. No.: 436,075

[22] Filed: May 8, 1995

[51] Int. Cl.<sup>6</sup> ..... A61N 5/00

[52] U.S. Cl. .... 600/7; 600/3

[58] Field of Search ..... 600/1-8

[56] References Cited

## U.S. PATENT DOCUMENTS

4,851,694 7/1989 Rague et al. .... 600/3  
5,030,194 7/1991 Van't Hooft ..... 600/7  
5,084,001 1/1992 Van't Hooft et al. .... 600/3  
5,092,834 3/1992 Bradshaw et al. .... 600/7

## FOREIGN PATENT DOCUMENTS

3335438 4/1985 Germany ..... 600/12

41 Claims, 7 Drawing Sheets

An afterloader for use in radiation oncology or intravascular radiotherapy comprises a wire storage mechanism for storing a radioactive-tipped source wire, a drive mechanism for advancing the source wire into a catheter or other channel implanted in a patient. A computer control system receives information from an encoder and a force sensor, which monitor the displacement of and the force exerted on the source wire, respectively. The displacement and force information are used to advance the wire at the fastest possible speed to the treatment site without exceeding a pre-programmed force profile designed to ensure the source wire does not puncture the catheter. The force profile is dynamic depending on the particular catheter being used, and catheter information may be inputted into the computer controller using a bar code or other information storage means in the catheter itself.

